

DIRECT MICROSCOPIC SOMATIC CELL COUNT
(Raw Commingled Cow, Goat, Sheep, Water Buffalo)
[Unless otherwise stated all tolerances are ±5%]

SAMPLES

1. **Laboratory Requirements (See CP, item 33 & 34)** _____

APPARATUS

2. **See Cultural Procedures, items 1-4** _____

a. Functional fume hood, face velocity 100 ft/min _____

1. Checked annually, records maintained, unit tagged _____

3. **Microscope Slides, Clean (see item 18), 2.54 x 7.62 cm** _____

a. 11.28 mm diameter areas delineated _____

b. Optionally, with center marks on sides of delineated area _____

c. Optionally, 5.08 x 7.62 or 5.08 x 11.43 cm with 11.28 cm delineated areas _____

4. **Pipetting Apparatus** _____

a. Metal Syringe (_____)

1. Suitable for rapid and convenient transfer of 0.01 mL of milk _____

2. Calibrated as specified in CP item 6e to deliver 0.0103 ±0.0005 g (average of 10 consecutive weighings with milk) _____

Avg. Wt. _____ Date ___/___/___ _____

3. Syringe etched with identification (imprinted serial number acceptable) and tagged with calibration date _____

b. Micropipettor, with appropriate tips (_____)

1. Suitable for rapid and convenient transfer of 0.01 mL of milk _____

2. Calibrated as specified in CP item 6e to deliver 0.0103 ±0.0005g (average of 10 consecutive weighings with milk) _____

Avg. Wt. _____ Date ___/___/___ _____

3. Micropipettor etched with identification (imprinted serial number acceptable) and tagged with calibration date

c. Records of calibration maintained

5. Dissecting Needle, Bent Point

a. Suitable for spreading milk film

6. Drying Device, Slide Drier or Incubator

a. Clean, dust-free, level surface

b. Heat source regulated at 40-45C

1. Temperature monitored with thermometer

7. Forceps or Slide Holder

a. Required for dipping and holding slides

8. Staining Jars or Trays

a. With tight fitting covers

b. Convenient size for holding solvents and stains

9. Slide Storage

a. Clean, dust-free insect-proof boxes, cases or files

10. Microscope Type: _____

a. Binocular with 1.8 mm oil immersion objective, rack and pinion sub-stage, condenser with iris diaphragm

b. Oculars, 10X (12X or 12.5X), Huygenian or wide-field

c. Optics provide a Single Strip Factor of 6070 or smaller

1. Each analyst measures field diameter and calculates SSF annually, round to three significant figures

2. Calculation of Single Strip Factor

a. Using a stage micrometer (item 11), measure field diameter (D) of oil immersion objective lens in mm

D = _____ mm

b. Compute SSF with formula: _____

$$SSF = 10,000 / (11.28 \times D)$$

SSF is _____

d. Mechanical Stage _____

1. Suitable for examination of slides, smooth action, does not drift, allows proper tracking of smears _____

e. Microscope Lamp, provides adequate illumination _____

11. Stage Micrometer Ruled with 0.1 and 0.01 mm Divisions _____

12. Hand Tally, accurate _____

MATERIALS

13. Immersion Oil _____

a. Refractive index 1.51-1.52 at 20C _____

14. Levowitz-Weber Modification of the Newman-Lampert Stain _____

a. Slowly add 0.6 g certified methylene blue chloride to 52 mL of 95% ethyl alcohol and 44 mL of tetrachloroethane (reagent grade) in a 200 mL flask and swirl to dissolve _____

b. When making stain, use gloves and prepare in fume hood (tetrachloroethane is TOXIC) _____

c. Let stand for 12-24 hr at 4.4-7.2C _____

d. Filter through Whatman No. 42 filter paper or equivalent _____

e. Add 4 mL of glacial acetic acid _____

f. Store in a clean, tightly closed container (traces of water or solvent may cause problems with this stain) _____

g. Or, Commercially prepared (xylene or tetrachloroethane) _____

Brand _____ Lot No _____ Exp Date _____

15. Canadian Formula Stain _____

a. Commercially prepared (xylene or tetrachloroethane) _____

Brand _____ Lot No _____ Exp Date _____

16. Alternate Methylene Blue Stain

a. Prepare as in item 14 with reagents:

- 1. Combine: 0.5 g cert. methylene blue chloride
56 mL 95% ethyl alcohol
40 mL xylene
4 mL glacial acetic acid

17. Pyronin Y-Methyl Green Stain for Goat or Sheep Milk

a. Carnoy's fixative

- 1. Combine: 60 mL chloroform
20 mL glacial acetic acid
120 mL 100% ethyl alcohol

2. Or, Commercially Prepared

Brand _____ Lot No _____ Exp Date _____

b. Pyronin Y-methyl green stain

- 1. Combine: 1.0 g Pyronin Y
0.56 g methyl green
196 mL water

2. Filter through Whatman No. 1 paper before use

3. Stain is light sensitive; store in brown bottle

4. Or, Commercially Prepared

Brand _____ Lot No _____ Exp Date _____

18. Slides, Cleaning

a. Physically clean

b. New slides may be cleaned by soaking in strong cleaning solution

c. Rinse thoroughly in flowing water 10-15 sec and MS water

d. Used slides may be soaked in hot detergent or wetting agent until all residues are removed, rinsed as above

e. Air or heat dry with minimal exposure to dust, insects, etc. and store dry

f. Or, store slides in alcohol and flame just before use

PROCEDURE

19. Slide Identification

- a. Legibly and indelibly identify each sample area on margin of slide

20. Sample Agitation

- a. Mix samples by shaking 25 times in 7 sec with 1 ft movement, sample removed within 3 minutes
- b. Optional: Warm high fat samples to 40C for no longer than 10 minutes prior to testing (discard after testing)

21. Sample Measurement and Smear Preparation (Metal Syringe)

- a. Before use and between successive samples, rinse syringe 2 - 3 times in clean, 25-35C water
- b. Before transferring test portion to slide, insert syringe not over 1 cm below surface (avoiding foam) of milk and repeatedly rinse
- c. Holding tip beneath surface, rinse syringe three times with milk, then fully depress and release plunger and withdraw test portion
- d. With clean paper tissue, remove excess milk from exterior of tip (with syringe tip up, wipe downward away from tip)
- e. Holding instrument vertical, place tip near center of area for smear, touch the slide with the tip and expel the test portion
1. With plunger still fully depressed, touch off once against a dry spot
 2. Do not release plunger until after touching off and removing tip from slide
 3. Spread milk with point of bent needle point (item 5), not hockey stick style
 4. Wipe needle dry between samples on tissue
- f. After spreading test portion, dry smears at 40-45C within 5 min on level surface (see item 6)
- g. To prevent smears from cracking and peeling from slide during staining, do not heat too rapidly
- h. Protect smears and slides from damage until read

22. Metal Syringe Cleaning

- a. Do not allow residues to dry on instrument
- b. Immediately after use, carefully disassemble and clean syringe
- c. Do not remove spring unless necessary
- d. Use only soap-less detergents and/or fat solvents sparingly as needed
- e. Clean all residues from measuring tube circulating detergent with bulb on delivery end
- f. Clean piston with dry paper tissue

23. Sample Measurement and Smear Preparation (Micropipettor)

- a. Use new tip for each sample
- b. Depress plunger and insert tip below surface (avoiding foam), fully release plunger slowly, remove tip from sample
- c. If necessary, remove excess milk from exterior of tip by wiping away from the tip with clean paper tissue
- d. Holding instrument vertical, place tip near center of area for smear, expel test portion
- e. Spread milk with point of bent needle point (item 5), not hockey stick style
- f. Wipe needle dry between samples on tissue
- g. After spreading test portion, dry smears at 40-45C within 5 min on level surface (see item 6)
- h. To prevent smears from cracking and peeling from slide during staining, do not heat too rapidly
- i. Protect smears and slides from damage until read

24. Staining Films

- a. Levowitz-Weber and Methylene Blue Stains
 - 1. Use ventilated hood for steps 2-4
 - 2. Submerge or flood slides with fixed, dried smears in stain for 2 min (timer used)

3. Drain off excess stain by resting edge of slide on absorbent paper _____
4. Dry thoroughly (air dry or use cool forced air) _____
5. Dip dry stained slides in 3 changes of tap water at 35-45C _____
6. Drain and air dry slides before examining smears _____

b. Pyronin Y-Methyl Green Stain (New York Modification)
Note: Stain is light sensitive and must be protected from overexposure to light. _____

1. Slide is run through the following staining scheme _____

Carnoy's fixative 5 min
50% Ethanol 1 min
30% Ethanol 1 min
H2O 1 min
Stain 6 min
N-Butyl alcohol flush briefly
Xylene flush briefly _____

- a. Optionally, if smears will not adhere to slides: _____

1. Allow slide to dry, (approx 10 min) protected from overexposure to light, after Carnoy's fixative step but before the 50% ethanol step OR _____
2. Allow slide to dry (approx 10 min) protected from overexposure to light, after stain step but before flushing with N-Butyl alcohol _____

2. Cells stain blue or blue-green; RNA and background stain pink _____

25. Examination _____

- a. Adjust microscope lamp to provide maximal optical resolution _____
- b. Locate edge of smear to be read using low power _____
- c. Place 1 drop immersion oil on smear _____
- d. Carefully lower oil immersion lens _____
- e. Focus and locate center of edge of area and begin counting cells _____
- f. Count all cells in field wide strip across diameter of a single smear, focusing up and down as necessary _____

- g. Identifying and counting somatic cells _____
 - 1. Cells possess a nucleus stained dark blue (bovine) or blue or blue-green (caprine) _____
 - 2. Cells generally 8 microns or larger (bovine; caprine may be smaller); do not count cells less than 4 microns; fragments counted only if more than 50% of nuclear material visible _____
 - 3. Cluster of cells counted as one unless nuclear units are clearly separated; focus up and down to ensure that there are no bridges connecting nuclear masses _____
 - 4. Count cells touching only top or bottom half of strip _____
 - 5. If in doubt, do not count _____
- h. After examination of each smear record strip count _____
- i. Conduct monthly comparative counting between analysts (refer to SPC item 19) _____

26. Slide Storage _____

- a. Remove oil by dipping in xylene (or equivalent), 15-20 sec _____
- b. Air dry _____
- c. Place in suitable storage (item 9) _____

REPORTS

27. Records and Reporting _____

- a. Maintain record of strip count for each smear examined _____
- b. Compute DMSCC/mL, multiply number of cells counted (strip count) by the SSF (item 10.c.2.b.) _____
- c. Report somatic cell counts as DMSCC/mL, record only first two left hand digits, round as necessary _____
 - 1. If the third digit is 5 round the second number using the following rules _____
 - a. When the second digit is odd round up (odd up, 235 to 240) _____
 - b. When the second digit is even round down (even down, 225 to 220) _____